

Inspection protocol

Product: **EX-TEC® PM 400**

Serial no.:

--	--	--

Test gas: **ExTox CAT**

e. g.: 050 30 0501

Gas type: Methane

Date								
	day							
	month							
	year							

1 Device condition								
1.1	Perfect condition ¹							
1.2	Fine dust filter OK ¹							
1.3	Disposable/rechargeable battery charge ²							

2 Pump check								
2.1	Pump error F100 in seal ¹							

3 Warning								
5.1	Zero point Sensitivity O ₂ Test gas: clean air							
	CH ₄ (Target: 0 % LEL/% vol., permissible: -3 – 3 % LEL or -0.15 – 0.15 % vol.)							
	CO (Target: 0 ppm, permissible: -3 – 3 ppm)							
	CO ₂ (Target: 0.04 % vol., permissible: -0.02 – 0.10 % vol.)							
	O ₂ (Target: 20.9 % vol., permissible: 19.9 – 21.9 % vol.)							
3.2	Sensitivity Test gas: ExTox CAT							
	CH ₄ (Target: 2.20 % vol., permissible: 45 - 55 % LEL or 2.00 – 2.40 % vol.)							
	CO (Target: 40 ppm, permissible: 36 – 44 ppm)							
	CO ₂ (Target: 2.00 % vol., permissible: 1.80 – 2.20 % vol.)							
	O ₂ (Target: 17.5 % vol., permissible: 16.5 – 18.5 % vol.)							
3.3	Visual alarm ¹							
3.4	Audible alarm ¹							

4 Comments								
	For example: – Housing damaged – Filter changed – Servicing necessary – ...							

Signature								

¹ For example: **Y** for 'Yes', **N** for 'No'

² For example: ¼

Enter current measurement values (actual values) at zero point and sensitivity.