Inspection protocol	***EX-TEC® SNOOPER 4*** **(methane CH_4)**	
Sensor:	PPM	
Serial no. (e. g.: 061 10 0001)		

SEWERIN
Protecting Water, Gas and Life.

24.02.2009

1.0	**Device status**										
1.1	- status correct (e.g.: Y / N)										
1.2	- remaining operating hours (e.g.: 5 h)										

2.0	**Pump test**										
2.1	- fault F100 when sealing										
2.2	- pump filter changed (e.g.: Y / N)										

3.0	**PPM measuring range**										
3.1	zero point (fresh air) - display 0 ppm										
3.2	test gas (1.00 % vol. CH_4) - display 0.8 – 1.5 % vol.										

4.0	**AL1 alarm triggering**										
4.1	optical alarm (e.g.: Y / N)										
4.2	audible alarm (e.g.: Y / N)										

5.0	**Observations** - housing broken - adjustment, repair - factory inspection - or the like										

6.0	**Test**										
	- day										
	- month										
	- year										
	- signature										