

Serial no. (e.g.: 066 11 0501)

--	--	--

04.12.2018

[illegible][illegible]

3.0	Fresh air								
3.1	Zero point CH ₄ – Display -1 – +1 % vol.								
3.2	Zero point CO ₂ (0.04 % vol.) – Display -1 – +1 % vol.								
3.3	Oxygen O ₂ (20.9 % vol.) – Display 20.4 – 21.4 % vol.								
3.4	Zero point H ₂ S – Display -10 – +10 ppm								

[illegible]

5.0	Carbon monoxide CO								
5.1	Zero point (fresh air) – Display -3 – +3 ppm								
5.2	Test gas 40 ppm – Display 37 – 43 ppm								

[illegible][illegible]