EU declaration of conformity

EX-TEC® PM 550/500

Portable battery-powered gas measuring device

Hermann Sewerin GmbH
Robert-Bosch-Str. 3
33334 Gütersloh – Germany

The product is in conformity with the relevant Union harmonisation legislation:

2011/65/EU

2014/30/EU

2014/34/EU

For evaluation of conformity the following harmonised standards apply:

EN 50270:2015	EN 60079-11
EN 60079-0	EN 61000-6-2:2005
EN 60079-1	EN 61000-6-4:2011
EN 60079-7	

For methane and propane, the following also apply for the Warning application:

EN 50271
EN 60079-29-1

The following certificates have been issued:

Notified body:	TÜV NORD CERT GmbH (0044)	DEKRA Testing and Certification GmbH (0158)
Certificate:	TÜV 17 ATEX 171969 X	BVS 19 ATEX G 002 X

This declaration of conformity is issued under the sole responsibility of Hermann Sewerin GmbH.

Signed for and on behalf of:
Gütersloh, 2023-06-16

Dr. Swen Sewerin (General Manager)